

Cellufine ETclean

Endotoxin inactivation of laboratory hardware and reagents by dry heat sterilization¹⁾

Table 1. Preparation of laboratory hardware and reagents used for Cellufine ETclean evaluation

Hardware & Chemicals	Operation
Glassware, metal, aluminium foil	Dry heat sterilization 30 minutes or more at 250 °C.
Reusable screw caps ²⁾ Red PBT (polybutylene terephthalate) Red melamine ETFE pouring rings	Dry heat sterilization 2 hours at 180 ° C
*Silicon / silicon tubes	Dry heat sterilization 2 hours at 180 °C.
Plastics	5% hydrogen peroxide 3 hours at 70 °C or 0.2 M NaOH/95 % EtOH 3 hours at room temperature.
**NaCl	Dry heat sterilization 30 minutes or more at 250 °C.
NaOH/HCl/Acetic acid	Can be used without treatment
Water	Pyrogen free water, water for injection
Buffers	ET-free reagent can be purchased or Cellufine ETclean column is used to make buffer Endotoxin-free

*Please carry out with reference to the description of a silicone product.

**If mineral salt is below the decomposition temperature, it can be dry heat sterilization.

References

1) APPLIED AND ENVIRONMENTAL MICROBIOLOGY, Nov. 1978, p. 710-714 Vol. 36, (5)

Dry-Heat Destruction of Lipopolysaccharide: Dry-Heat Destruction Kinetics

KIYOSHI TSUJI AND SUSAN J. HARRISON

2) <http://www.sigmaaldrich.com/>

JNC CORPORATION

2-1, Otemachi 2-chome, Chiyoda-ku

Tokyo 100-8105 Japan

cellufine@jnc-corp.co.jp