

Operating Instructions

Strong cation exchange Chromatography Media Cellufine™ MAX GS

Description

Cellufine MAX is the new name of 2nd generation Cellufine chromatography media. Cellufine MAX GS strong cation exchanger is highly cross-linked media with stable at high flow velocities and with optimized ligand density. These high performances offer significant opportunity for increasing downstream purification throughput, especially aggregate removal from therapeutic Mabs.

Characteristics of Cellufine MAX GS

Type	
Ion exchange type	Strong cation (graft polymer type)
Matrix	Highly cross-linked cellulose
Particle size	ca. 40 – 130 μm (Average 90)
Ion exchange capacity	0.09~0.15 m mol /ml
Flow velocity	600 cm/h(0.3 MPa) I.D.30 cm-L 20 cm, pure water at 24 °C
Adsorption capacity	≥ 100 mg / ml (Lysozyme)
Dynamic binding capacity	≥ 70 mg /ml (Polyclonal IgG 10 % DBC)
pH stability (20°C, 1week)	2 – 13
Chemical stability	Stable all commonly used aqueous buffers, 0.5 M NaOH
Storage	20 % ethanol

Packing to the column

- Calculate volume required for the desired bed dimension.
 - Packed bed volume = column cross sectional area (cm²) x bed height (cm)
 - Required sedimented gel volume = Packed bed volume x 1.1
 - The slurry concentration in the bottle of Cellufine MAX GS is approximately 50% in 20% ethanol.
- Wash the gel with water or the appropriate buffer.
- Prepare a 40 – 60 % (v/v) slurry with water, 0.1 M NaCl solution or the appropriate buffer. Equilibrate at ambient temperature for one hour.
- Gently stir. If required place under vacuum to degas.
- Column
 - A column is prepared according to the instructions from the column supplier.
 - The bed support should be wetted in a packing solution or 20% ethanol before use to remove air.
 - Pour the packing solution into the column tube and it check that solution flows out from the column exit. Shut the exit valve when approximately 0.5 to 1cm height of the solution remains.
- Carefully pour the slurry into the column without creating air bubbles. Depending on the volume, a filler tube may be necessary.
- Mount the top adapter on the top of column. (Be careful to not entrain air)
- Open the column outlet and begin pumping elution buffer for 10 min at 1000cm/h or 0.3 MPa. Caution: do not excess the operation pressure limit for the selected column.

9. Mark the gel bed height. Stop the pump and shut the column outlet valve.
10. Disconnect the top adapter line from the pump. Loosen the adaptor seal and move the top adapter down to mark the gel bed reached at packing.
11. After the bed stabilizes, lock the adapter and reconnect the line from the pump. Equilibrate with 10 column volumes of adsorption buffer before sample loading.

Packing to the fixed length column

1. Calculate volume required of the desired bed dimension.
 - (a) Packed bed volume= column cross sectional area (cm²) x bed height (cm)
 - (b) Required sedimented gel volume=Packed bed volume x 1.1
 - (c) Note: When using a packing connector, the amount of excess gel corresponds to the required sedimentary gel volume.
 - (d) The slurry concentration in the bottle of Cellufine MAX GS is approximately 50 % in 20 % ethanol.
2. Wash the gel with water or the appropriate buffer.
3. Prepare 40– 60 % (v/v) slurry with water or the appropriate buffer (high salt). Equilibrate at ambient temperature for one hour.
4. Gently stir. If required place under vacuum to degas.
5. Column
 - (a) A column is prepared according to the instructions from the column supplier.
 - (b) The bed support wetted in a packing solution or 20 % ethanol before use to remove air.
 - (c) Pour the packing solution into the column tube and it check that solution flows out from the column exit. Shut the exit valve when approximately 0.5 to 1cm high of the solution remains.
6. Carefully pour the slurry into the column and packing connector without creating air bubbles. Depending on the volume, a filler tube may be necessary.
7. Mount the top adapter on the top of packing connector.
8. Open the column outlet and begin pumping elution buffer for 10 min at 1000cm/h or 0.3Mpa. Caution: do not excess the operation pressure limit for the selected column.
9. Stop the pump and shut the column outlet valve.
10. Disconnect the top adapter line from the pump. Remove the packing connector. Before remove excess medium from the packing connector, if necessary.
11. Mount the top adapter, lock the adapter and reconnect the line from the pump. Equilibrate with 10 column volumes of adsorption buffer before sample loading.

Evaluation of packing

See appendix 1

Operating Guidelines

General Operation

Typically, adsorption to Cellufine MAX Cation Exchange medium occurs in relatively low ionic strength (e.g., below 0.1 M NaCl) in the pH range from 4.0- 9.0 below the *pI* of the target protein.

The binding capacity is strongly affected by pH and conductivity. Under these conditions, proteins with neutral or net negative charge will bind. Bound components are then resolved by stepwise elution with buffers containing progressively higher salt concentration or by elution with a linear salt gradient.

Sample Preparation and Load

Prepare sample by centrifugation or microfiltration to remove insoluble material. Samples should be prepared to a concentration of 1 – 20 mg/ml, in binding buffer or at comparable

conditions of ionic strength and pH. If necessary, exchange sample buffer using dialysis, diafiltration or desalting chromatography.

Recommended Buffers

Adsorption buffer: 0.01–0.05 M sodium acetate (pH 4 to 6) or 0.01-0.05 M phosphate or Tris-HCl (pH 6 to 9)

Elution buffer: 0.1 – 2.0 M sodium chloride in adsorption buffer.

Other common buffer systems may be used. For additional information on chromatographic methods of protein purification, see References.

Regeneration and equilibration

After separation, wash the binding material with over 5-bed volumes of high ionic strength solution (1-2 M NaCl). After washing the column, pump an additional 5-bed volume of adsorption buffer, or until the column eluate will be stable pH and conductivity values.

Depyrogenation

Wash the column with 5-bed volumes of 0.2 M NaOH, let stand for 16 hours, and then wash with endotoxin-free water or equilibration buffer.

• 0.2 M NaOH-20% EtOH is effective for endotoxin removal. Moreover 0.2 M NaOH-90% EtOH can rapidly decrease LPS, with contact for at least 2 hours.

Chemical and Physical Stability

Stable in:

Most salts (NaCl, (NH₄)₂SO₄, etc.), Alcohol (30 % (v/v) IPA, 70 % (v/v) EtOH), Urea (6 M) and Guanidine-HCl (6 M)

pH 2-13 at 20 °C, 1 week

Cleaning-in-place (CIP)

Cellufine MAX GS performance remains constant for at least 100 CIP operating cycles employing 0.5 M NaOH by 10 CV washing.

Flow Rate

Cellufine MAX GS is based on highly cross-linked cellulose gel, and are stable at high flow velocities.

1,000 cm/h in a 2.2 cm diameter column with 20cm bed height at < 0.3 MPa.

Over 500 cm/h 30 cm diameter column with 20 cm bed height at < .3 MPa.

Storage

Store unopened container at ambient temperature. Do not freeze.

Short-term storage for bulk and column (2 weeks or less) can be at a room temperature with pH 2 to pH 13. It is possible to store under alkaline condition at less than 20 °C as recommendation.

0.5 M NaOH ; for less than 2 weeks, at 20 °C

Longer storage should be in neutral buffer containing 20 % ethanol, at 2 – 25 °C. Do not freeze.

Shelf Lifetime

5 years from date of manufacture

References

1. Harris, E.L.V. and Angal, S., *Protein Purification Methods: A practical Approach*. New York: Oxford University Press, 1989.
2. Janson, J.-C. and Ryden, L., *Protein Purification: Principles, High Resolution Methods, and Applications*. 2nd ed. New York: John Wiley & Sons, Inc., 1998

Product Ordering Information (Catalogue No.)

Media type	Pack Size					
	MC* 1mL x 5	MC 5mL x 5	100 mL	500 mL	5 L	10 L
Cellufine™ MAX GS	21300-51	21300-55	21300	21301	21302	21303

MC = Mini-Column

JNC CORPORATION

Life Chemicals Division

2-1, Otemachi 2-Chome, Chiyoda-ku, Tokyo 100-8105, Japan

Phone +81-3-3243-6150, Fax+81-3-3234-6219

E-mail: cellufine@jnc-corp.co.jp


<http://www.jnc-corp.co.jp/fine/en/cellufine>

Appendix 1: Evaluation of column packing of Cellufine

Evaluation methods for the packing status of a column can employ indices, such as the number of (theoretical) plates (N), the height equivalents to a theoretical plate (HETP), and a symmetry factor (As).

The selected evaluation indices are affected by the measurement condition. For example, they vary with changes in the difference between the diameter/height of a column, the difference in piping, solvent sample volume, the flow velocity, temperature, etc. Therefore, it is necessary to use the same measurement conditions each time.

Conditions	
Sample volume	1 % (MAX 2.5%) of column bed volume
Sample concentration	1-2 % (V/V) acetone or 1 M NaCl
Flow rate	~30 cm/h (X mL/hr/column cross section)
Detector	UV, conductivity



Formula
HETP = L/N
$N = 5.54 \times (V_e/W_h)^2$
As = b/a

L	Column length [cm or m]
V_e	Elution time or volume
W_h	Half of width of peak
a,b	Peak width of 10% peak height (a)front, (b)rear
Note	V_e, W_h and a, b should have same dimensional units

(Note)

Generally, a larger value of N is good. (Likewise a smaller value of HETP is good.) The asymmetry factor value (As) should be close to 1. Generally, acceptable symmetry values range from 0.8-1.6.

TERMS AND CONDITIONS OF SALE

1. **Acceptance**-Buyer's placement of this order shall create a contract subject to and expressly limited by these terms and conditions. Acceptance may only be made on the exact terms and conditions hereof and if additional or different terms are proposed by Buyer, such response shall constitute a counteroffer. **THE TERMS OF THIS CONTRACT SHALL SUPERSEDE ANY CONFLICTING TERMS CONTAINED ON BUYERS PURCHASE ORDER OR ANY DOCUMENT OR INSTRUMENT SUBMITTED BY BUYER.**

2. **Prices, Taxes and Payment** - All prices are firm unless otherwise agreed to in writing. JNC Corporation reserves the right to change the prices and specifications of its Products at any time without notice. Any tax, duty, custom or other fee of any nature imposed upon this transaction by any federal, state or local governmental authority shall be paid by Buyer in addition to the price quoted or invoiced. In the event JNC Corporation is required to prepay any such tax, Buyer will reimburse JNC Corporation. Payment terms shall be net 30 days after shipment by JNC Corporation. An interest charge equal to 1½% per month (18% per year) will be added to invoices outstanding beyond 30 days after shipment. In addition JNC Corporation reserves the right to require C.O.D. payment terms from any Buyer whose account is overdue for a period of more than 60 days or who has an unsatisfactory credit or payment record. JNC Corporation may also refuse to sell to any person until overdue accounts are paid in full.

3. **Delivery and shipment** - JNC Corporation will make every effort to ship the Products or provide the services hereunder in accordance with the requested delivery date, provided, that JNC Corporation accepts no liability for any losses or for general, special or consequential damages arising out of delays in delivery. Shipment of all Products shall be F.O.B. point of distribution by JNC Corporation; identification of the Products shall occur when they leave JNC Corporation's point of distribution, at which time title and risk of loss shall pass to Buyer. All shipment costs shall be paid by Buyer and if prepaid by JNC Corporation the amount thereof shall be reimbursed to JNC Corporation.

4. **Inspection** - Buyer shall be responsible for inspecting all Products shipped hereunder prior to acceptance, provided, that if, Buyer shall not have given JNC Corporation written notice of rejection within 30 days following shipment to Buyer, the Products shall be deemed to have been accepted by Buyer.

5. **Disclaimer of Express and Implied Warranties** - The Products shall be covered by the applicable JNC Corporation standard warranty. **NO OTHER EXPRESS OR IMPLIED WARRANTY IS MADE WITH RESPECT TO THE PRODUCTS. JNC CORPORATION EXPRESSLY EXCLUDES THE IMPLIED WARRANTIES OF MERCHANTABILITY AND OF FITNESS FOR A PARTICULAR PURPOSE.** Any model or sample furnished to the Buyer is merely illustrative of the general type and quality of goods and does not represent that the Products will conform to the model or sample. Buyer's remedies under JNC Corporation's warranty shall be limited to repair or replacement of the Product or component which failed to conform to JNC Corporation's warranty. **JNC Corporation shall not be liable for any consequential damages or economic loss or property damage incurred by Buyer.**

6. **Returned Goods** - No Products shipped under this contract may be returned without the express prior authorization of JNC Corporation. All returns of Products are subject to a restocking charge. No returns will be authorized after 120 days following shipment to Buyer.

7. **Technical Advice** - JNC Corporation may, at Buyer's request, furnish technical assistance, advice and information with respect to the Products if and to the extent that such advice, assistance and information is conveniently available. It is expressly agreed that there is no obligation to

provide such information which is provided without charge at the Buyer's risk and which **is provided subject to the disclaimers set forth in paragraph 5 above.**

8. **Agents, etc.** - No agent, employee or other representative has the right to modify or expand JNC Corporation's standard warranty applicable to the Products or to make any representations as to the Products other than those set forth in JNC Corporation's product literature and any such affirmation, representation or warranty, if made, should not be relied upon by Buyer and shall not form a part of this contract.

9. **Fair Labor Standards** - JNC Corporation represents that the Products or services provided hereunder were produced and/or performed in compliance with the requirements of all sections of the Fair Labor Standard Act of 1938, as amended.

10. **Equal Employment Opportunity** - JNC Corporation is an Equal Opportunity Employer. It does not discriminate in any phase of the employment process against any person because of race, color, creed religion, national origin, sex, age, veteran or handicapped status. The JNC Corporation Equal Opportunity Certificate, which is mailed annually to all vendors and vendees, is incorporated into this contract by reference.

11. **Modifications, Waiver, Termination** - This contract may be modified and any breach hereunder may be waived only by a writing signed by the party against whom enforcement thereof is sought.

12. **Governing Law** - This contract shall be governed by and construed in accordance with the laws (other than those relating to conflict of laws questions) of the Commonwealth of Japan.

13. **Arbitration** - Any and all disputes or controversies arising under, out of or in connection with this contract or the sale or performance of the Products shall be resolved by final and binding arbitration in Tokyo under the rules of the Japan Arbitration Association then obtaining. The arbitrators shall have no power to add to, subtract from or modify any of the terms or conditions of this contract. Any award rendered in such arbitration may be enforced by either party in either the courts of the Commonwealth of Japan District Court for the District of Japan, to whose jurisdiction for such purposes JNC Corporation and Buyer each hereby irrevocably consents and submits.